

ANESTHESIA MACHINE WITH VENTILATOR RE 907-C7



Specification

- ◆ Respiration mode: IPPV, MANUAL
- ◆ Tidal Volume: 50~1600ml
- ◆ Respiratory Frequency: 1~90bpm
- ◆ I:E ratio: 4:1~1:6
- ◆ Working Mode: Semi~open, Semi~closed or Closed system
- ◆ Oxygen flush: 25L/min~75L/min
- ◆ High light LED display
- ◆ Anesthesia Vaporizer: One Isoflurane
- ◆ Anesthesia Vaporizer: 120ML
- ◆ Flowmeter: 2 Tube
- ◆ Mains: 100~240V, 50/60Hz
- ◆ Alarm parameters
- ◆ Power consumption: ≤50W



ANESTHESIA MACHINE WITH VENTILATOR RE 907-C7 IBIS200



Specification

- ◆ Respiration mode: IPPV, MANUAL
- ◆ Tidal Volume: 50~1600ml
- ◆ Respiratory Frequency: 1~90bpm
- ◆ I:E ratio: 4:1~1:6
- ◆ Working Mode: Semi~open, Semi~closed or Closed system
- ◆ Oxygen flush: 25L/min~75L/min
- ◆ High light LED display
- ◆ Anesthesia Vaporizer: One Isoflurane IBIS 200
- ◆ Anesthesia Vaporizer: 250ML
- ◆ Flowmeter: 2 Tube
- ◆ Mains: 100~240V, 50/60Hz
- ◆ Alarm parameters
- ◆ Power consumption: ≤50W

